

Uncut sheet of COVID-19 antigen detection

Certificate of Analysis

CAT. NO.: 2011K401

Product Name	Uncut sheet of COVID-19 antigen detection Novel coronavirus (SARS-CoV-2) antigen detection kit (colloidal gold immunochromatography)- detection membrane		
Spec.	60mm*300mm	Lot No.	2011K401
Production quantity	612 pieces	Sampling Amount	8 pieces
Production Date	2020-11-01	Exp.	2023-04
Store at	sealed at 2~30°C	Reported	2020-11-02
Inspection record			
Test items	Acceptance Criteria	Result	Conclusion
Appearance	Clean, no scratches, smooth surface, no damage, no burrs, no stains; Each support of the large board should be firmly pasted.	<input checked="" type="checkbox"/> Meet the requirements <input type="checkbox"/> non-compliant	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Length (mm)	300mm±3mm	300mm	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
C line, T line position	T line, B position should be 30mm±0.5mm; C line B position should be 34.5mm±0.5mm	T line, B position is: 30.0mm C line, B position is: 34.5mm	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Liquid moving speed (mm/min)	≥10mm/min	57.5mm/min	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Fading time of film (min)	15~20min	16min	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Positive rate of positive reference	Use the corporate positive reference product P1-P5 to test, The result should meet the requirement as follows: The color intensity of P1 should be C8-C7;	<input checked="" type="checkbox"/> Meet the requirements <input type="checkbox"/> non-compliant	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified



	<p>The color intensity of P2 should be C6~C5;</p> <p>The color intensity of P3 should be C5~C4;</p> <p>The color intensity of P4 should be C8~C7;</p> <p>The color intensity of P5 should be C7~C6.</p>		
Minimum detection limit	<p>Use corporate lowest detection limit reference product L1~L3 for testing, the result should meet the requirement as follows:</p> <p>The color intensity of L1 should be B;</p> <p>The color intensity of L2 should be C9~C7;</p> <p>The color intensity of L3 should be C8~C7.</p>	<input checked="" type="checkbox"/> Meet the requirements <input type="checkbox"/> non-compliant	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Precision	<p>Use corporate repeatable reference product R1~R2 to test, and the result should meet the requirement as follows:</p> <p>The color intensity of R1 should be C6~C5;</p> <p>The color intensity of R2 should be C7~C6.</p>	<input checked="" type="checkbox"/> Meet the requirements <input type="checkbox"/> non-compliant	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Negative rate of Negative reference	<p>Test corporate negative reference product N1~N7, all color intensity should be B.</p>	<input checked="" type="checkbox"/> Meet the requirements <input type="checkbox"/> non-compliant	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Random Clinical Sample	<p>Sampling 100 samples randomly, using the sample extraction solution to elute the sample and test, The test color intensity should be B~C9.</p> <p>Sampling 20 samples randomly in UTM tube with COPAN preserved solution, and test, The test color intensity should be B~C9.</p>	<p>The number of color intensity of B~C9/total number of tests: <u>120 / 120</u></p>	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
	<p>The total false positive rate should be $\leq 2\%$</p>	<p>Total false positive rate: <u>0</u></p>	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified
Remark	/		
Conclusion:	<input checked="" type="checkbox"/> qualified <input type="checkbox"/> Unqualified		